



MicroMedia Trading House
support@micromediatradinghouse.com

Certificate of Analysis

Name: CLED DI Medium
Product code: MM0096
Lot Number: 16240105
Manufacture Date: JUN 2016
Expiry Date: JUN 2020

Dehydrated Media:

Colour: Light beige
Particle size: Fine, homogenous
Solubility: 1-2 min.

Prepared Medium:

Directions: 37,1 g/l
Clarity: Green/blue clear gel
Deposit: None
Gel strength: Correct
pH: 7.5±0.2
Sterilization: By autoclaving at 121 C• for 15 minutes
Remarks:

Microbiological Response:

| | Result | Colonies |
|------------------------------|--------|---------------|
| E. coli ATCC25922 | growth | red |
| S. aureus ATCC25923 | growth | yellow |
| Proteus mirabilis ATCC 12453 | growth | greenish blue |

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date: 15 September 2016

This certificate was generated by computer and does not require any signature.