



MicroMedia Trading House  
support@micromediatradinghouse.com

## Certificate of Analysis

**Name:** CLED SI Medium  
Product code: MM0097  
Batch Number: 1314308  
Manufacture Date: APR 2013  
Expiry Date: APR 2017

**Dehydrated Media:**

Colour: Light beige  
Particle size: Fine, homogenous  
Solubility: 1-2 min.

**Prepared Medium:**

Directions:\* 37.05 g/l  
Clarity: Clear, green/blue  
Deposit: None  
pH:\* 7.3 ± 0.2  
Sterilization: By autoclaving at 121 C• for 15 minutes

**Microbiological response after 18-24 hours incubation at 37 °C:**

Organism	Result / Colour
<i>Escherichia coli</i> ATCC 25922	Growth/ Yellow
<i>Salmonella typhimurium</i> ATCC 14028	Growth/ blue
<i>Staphylococcus aureus</i> ATCC 25923	Growth/ Yellow
<i>Proteus mirabilis</i> ATCC 12453	Growth/ blue

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date: 15 September 2016

This certificate was generated by computer and does not require any signature.