

# Certificate of Analysis

**Name:** **Pseudomonas agar**  
Product code: MM0423  
Batch Number: 13400405  
Manufacture Date: OCT 2013  
Expiry Date: OCT 2017

## Dehydrated Media:

Colour: light beige  
Particle size: Fine, powder  
Solubility: 1-2min.

## Prepared Medium:

Directions:\* 48.4 g/l  
Clarity: clear amber-slightly yellow agar  
Deposit: None  
Gel strength: Correct  
pH:\* 7.1±0.2  
Sterilization: Autoclaving at 121 °C for 15 minutes  
Remarks: Add 10 g glycerol to the medium before sterilisation

## Microbiological Response:

Organism	CFU	Result	Colour
Staph. aureus ATCC 25923	100-1000	inhibited	Emerald
Pseudomonas aeruginosa ATCC 27853	100-1000	growth	yellowish-green pigmentation

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date: 16 September 2016