



MicroMedia Trading House

Certificate of Analysis

Name:

Product code:

Batch Number:

Manufacture Date:

Expiry Date:

Sensitivity Test Agar

MM0121

15200202

MAY 2015

MAY 2029

Dehydrated Media:

Colour:

whitish to
yellowish

Particle size:

Fine,
homogenous

Solubility:

1-2min.

Prepared Medium:

Directions:*

40,0g/l + 7% lysed horse blood

Clarity:

dependent upon the blood

Deposit:

None

Gel strength:

Correct

pH:*

7,4±0,2 (current batch 7,43)

Sterilization:

By autoclaving at 121°C for 15 minutes

Remarks:

Microbiological Reponse:

Organism	Growth
Escherichia coli ATCC 25922	good
Pseudomonas aeruginosa ATCC 27853	good
Staphylococcus aureus ATCC 25923	good

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date:19 September 2016