



MicroMedia Trading House
support@micromediatradinghouse.com

Certificate of Analysis

Name:	Iso-Sensitest Agar
Product code:	MM0085
Batch Number:	1039502
Manufacture Date:	SEPT 2010
Expiry Date:	SEPT 2014

Dehydrated Media:

Colour:	Light beige-straw coloured
Particle size:	Fine powdery
Solubility:	Good (0,5 min.)

Prepared Medium:

Directions:*	34,19 g/l
Clarity:	Hazy and yellow colour/ straw coloured gel
Deposit:	None
Gel strength:	Correct
pH:*	7.4±0.2
Sterilization:	By autoclaving at 121 C• for 15 minutes
Remarks:	

Microbiological Response:

Staphylococcus Aureus ATCC 25923
E. coli ATCC 25922
Enterococcus faecalis ATCC 29212
Pseudomonas aequinosa ATCC 27853

Result

growth
growth
growth
growth

The media conform to the specification of the NCCLS standard M2-A7.

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date:19 September 2016

This certificate was generated by computer and does not require any signature.