



Certificate of Analysis

Name: DNase Agar
Product code: MM0134
Batcher: 1047404_1
Manufacture Date: DEC 2010
Expiry Date: DEC 2014

Dehydrated Media:

Colour: Light beige
Particle size: Fine
Solubility: 1 min.

Prepared Medium:

Directions:* 40 g/l
Clarity: Light to medium, yellow to tan, clear to slightly hazy.
Deposit: None
Gel strength (where appropriate): Correct
pH:* 7,3±0,2
Sterilization: By autoclaving at 121°C for 15 minutes

Microbiological Response:

Organism	Result	Colour
Staphylococcus aureus ATCC 25923	Growth	Surrounded with a clear zone when flooded with 1N HCl
Staphylococcus epidermidis ATCC 14990	Partial inhibition	No clear zone around the colony
Serratia marcescens ATCC 8100	Positive	

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date: 02 February 2012