



**MicroMedia Trading House
HUNGARY**

Certificate of Analysis

Name: Urea Broth Base
Product Code: MM0028
LOT Number: 15420706
Manufacture Date: OCT 2015
Expiry Date: OCT 2020

Dehydrated Media:

Colour: cream
Particle size: Fine, powder
Solubility: 1-2min.

Prepared Medium:

Directions:* 9 g/l
Clarity: Clear, yellow broth
Deposit: None
pH:* 6.8-0,2 (cb : 6,85)
Sterilization: By autoclaving at 121°C for 15 minutes
Remarks: After autoclaving Cool to 50 °C and aseptically
add 50 ml of sterile Urea solution.
or 0,9 g to 95 ml of distilled water + 5 ml of
sterile urea solution

Microbiological Response:

Organism	Result urea reaction
<i>Proteus vulgaris</i> ATCC 13315	Positive
<i>Escherichia coli</i> ATCC 25922	Negative

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date of reprinting :26 September 2016