



**MicroMedia Trading House
HUNGARY**

Certificate of Analysis

Name: Urea Agar Base

Product code: MM0066
Batch Number: 14140303
Manufacture Date: MAR 2014
Expiry Date: MAR 2018

Dehydrated Media:

Colour: Off White
Particle size: Fine, powdery
Solubility: 1-2min.

Prepared Medium:

Directions:* 24,7 g/l
Clarity: Yellow/pale pink, translucent Complete medium is light to medium orange, clear to slightly hazy
Deposit: None
Gel strength: Correct
pH:* 6,8 • 0.2
Sterilization: Sterilise at 121 • C for 15 minutes
Remarks: Cool to 47 • C, add aseptically sterile urea solution

Econometric test:

	<u>Result</u>	<u>urease reaction</u>
Proteus vulgaris ATCC 8427	+	Positive
Salmonella choleraesuis subsp. choleraesuis serotype Typhimurium ATCC 13311	+	Negative
Escherichia coli ATCC 25922	+	Negative
Klebsiella pneumoniae ATCC 13883	+	Weak positive

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date: 26 September 2016

This certificate was generated by computer and does not require any signature.