

**Name:** **Maximal Recovery Diluent**

Product code: MM0029  
Batch Number: 13360105  
Manufacture Date: SEPT 2013  
Expiry Date: SEPT 2017

**Dehydrated Media:**

Colour: Cream to beige  
Particle size: Fine powdery  
Solubility: Good

**Prepared Medium:**

Directions:\* 9,6 g/l  
Clarity: Clear and pale yellow // Colorless  
Deposit: None  
Gel strength: Correct  
pH:\* 7.0 ±0,2  
Sterilization: By autoclaving at 121 C· for 15 minutes  
Remarks: To 10g test sample add 90ml MRD and mix well

**Microbiological Response:**

<b>Organism</b>	<b>Inoculum CFU</b>	<b>Result</b>
Staphylococcus aureus ATCC 25923	100-1000	No significant viable CFU reduction
E.coli ATCC 25922	100-1000	No significant viable CFU reduction

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

Date:03 October 2016